



DIVISION OF INDUSTRY SERVICES
Plumbing Product Review
P.O. Box 2658
Madison, Wisconsin 53701-2658
TTY: Contact Through Relay

Scott Walker, Governor
Dave Ross, Secretary

February 19, 2015

Revised Copy

SKYLINE MEDICAL INC.
DAN KUBES
2915 COMMERS DRIVE, STE 900
EAGAN MN 55121

SKYLINE MEDICAL INC.
DAVE DAUWALTER
2915 COMMERS DRIVE, STE 900
EAGAN MN 55121

Re: Description: HEALTH CARE PLUMBING APPLIANCE
Manufacturer: SKYLINE MEDICAL INC.
Product Name: STREAMWAY SYSTEM
Model Number(s): STM 2000
Product File No: 20150005

The specifications and/or plans for this plumbing product have been reviewed and determined to be in compliance with chapters SPS 382 through 384, Wisconsin Administrative Code, and Chapters 145 and 160, Wisconsin Statutes.

The Department hereby issues an approval based on the Wisconsin Statutes and the Wisconsin Administrative Code. This approval is valid until the end of February 2020.

This approval is contingent upon compliance with the following stipulation(s):

- Existing DWV piping must be large enough to handle the additional wastewater load generated by these devices.
- The wastewater from these devices shall be discharged to a sanitary drain, waste and vent (DWV) system comprised of materials that are stable when in contact with the maximum wastewater temperature generated by these devices.

No water supply cross connection control measures are required because there is no connection to the water supply system.

The department is in no way endorsing this product or any advertising, and is not responsible for any situation which may result from its use.

Sincerely,

Glen W. Schlueter
Environmental Engineer - Plumbing Product Reviewer
Department of Safety and Professional Services
Division of Industry Services
Bureau of Technical Services
(608) 267-1401 Phone
(608) 267-9723 Fax
glen.schlueter@wi.gov E-mail